



basic
laboratory

www.basiclab.com

2218 Railroad Avenue voice 530.243.7234
Redding, California 96001 fax 530.243.7494

3860 Morrow Lane, Suite F voice 530.894.8966
Chico, California 95928 fax 530.894.5143

June 07, 2012

Lab ID: 2050830

BETTY CREDIT

[REDACTED]
CHICO, CA 95926

RE: GENERAL TESTING [REDACTED] CHICO


Dear BETTY CREDIT ,

Enclosed are the analysis results for Work Order number 2050830. All analysis were performed under strict adherence to our established Quality Assurance Plan. Any abnormalities are listed in the qualifier section of this report.

If you have any questions regarding these results, please feel free to contact us at any time. We appreciate the opportunity to service your environmental testing needs.

Sincerely,


For



Ricky D. Jensen
Laboratory Director

California ELAP Certification Number 1677



basic
laboratory

www.basiclab.com

2218 Railroad Avenue voice 530.243.7234
Redding, California 96001 fax 530.243.7494

3860 Morrow Lane, Suite F voice 530.894.8966
Chico, California 95928 fax 530.894.5143

Report To: CREDIT, BETTY

[Redacted]
CHICO, CA 95926

Attention: BETTY CREDIT

Project: GENERAL TESTING [Redacted] CHICO

Lab No: 2050830

Reported: 06/07/12

Phone: [Redacted]
P.O. # [Redacted]

Metals - Total

Analyte	Units	Results	Qualifier	MDL	RL	Method	Analyzed	Prepared	Batch
WELL WATER RINSED OVER DIRTY SOLAR PANELS Water (2050830-01) Sampled:05/16/12 16:30 Received:05/17/12 10:34									
Aluminum	ug/l	52000		50	250	EPA 200.7	06/06/12	05/31/12	B2E0804
Barium	"	480		5	25	"	"	"	"
Strontium	"	413		5	25	"	"	"	"

Notes and Definitions

- DET Analyte DETECTED
- ND Analyte NOT DETECTED at or above the detection limit
- NR Not Reported
- dry Sample results reported on a dry weight basis
- RPD Relative Percent Difference
- < Less than reporting limit
- ≤ Less than or equal to reporting limit
- > Greater than reporting limit
- ≥ Greater than or equal to reporting limit
- MDL Method Detection Limit
- RL/ML Minimum Level of Quantitation
- MCL/AL Maximum Contaminant Level/Action Level
- mg/kg Results reported as wet weight
- TTLC Total Threshold Limit Concentration
- STLC Soluble Threshold Limit Concentration
- TCLP Toxicity Characteristic Leachate Procedure
- Note 1 Received Temperature - according to EPA guidelines, samples for most chemistry methods should be held at ≤6 degrees C after collection, including during transportation, unless the time from sampling to delivery is <2 hours. Regulating agencies may invalidate results if temperature requirements are not met.
- Note 2 According to 40 CFR Part 136 Table II, the following tests should be analyzed in the field within 15 minutes of sampling: pH, chlorine, dissolved oxygen, and sulfite.

Approved by
Basic Laboratory, Inc.
California ELAP Cert #1677 and #2718

BASIC LABORATORY CHAIN OF CUSTODY RECORD

3860 Morrow Lane, Suite F, Chico, CA 95928 (530) 894-8966 FAX (530) 894-5143

LAB #: **2050830**

CLIENT NAME: Betty Credit		PROJECT NAME:	PROJECT #:	PAGE 1 OF 1		
[REDACTED ADDRESS]		REPORT DUE DATE: 6-1-12	TURN AROUND TIME: Standard Rush	# OF SAMPLES: 1		
Chico, CA - 95926		ANALYSIS REQUESTED				
PROJECT MANAGER:						
PHONE: [REDACTED]	EMAIL: [REDACTED]	NUMBER OF BOTTLES	Aluminum	Barium	Strontium	
FAX:	RESULTS SENT: <input checked="" type="radio"/> Email <input type="radio"/> Fax <input type="radio"/> EDD <input checked="" type="radio"/> Mail					MATRIX / TYPE: W
INVOICE TO:	PO#:					CUSTODY SEAL INTACT? Yes No N/A
				SYSTEM #:		
				EDD TYPE:		
				QC: Standard Level II		
				LAB ID		
				CHLORINE RESIDUAL OR COMMENTS		

SAMPLE DATE	SAMPLE TIME	WATER	COMP	SOLID	SAMPLE LOCATION / IDENTIFICATION	Aluminum	Barium	Strontium	LAB ID	CHLORINE RESIDUAL OR COMMENTS
5/16/12	4:30	AM <input checked="" type="radio"/> PM <input checked="" type="radio"/>	<input checked="" type="checkbox"/>		[REDACTED] Chico, CA 95926	1	X	XX		
		AM PM			Reverse Osmosis					
		AM PM			Filtered well water					
		AM PM			Rinsed over dirty					
		AM PM			solar panels.					
		AM PM			(refridgerated)					
		AM PM			over night					
		AM PM								
		AM PM								
		AM PM								
		AM PM								
		AM PM								
		AM PM								
		AM PM								
		AM PM								
		AM PM								
		AM PM								
		AM PM								
		AM PM								
		AM PM								
		AM PM								
		AM PM								
		AM PM								

IP AT IT
 MAY 17 2012
 BY: **160 CASG**

5/17/12 @ 1035A
 ph L2

PRESERVED WITH: HNO₃ H₂SO₄ NaOH ZnAc₂/NaOH HCL NaThio OTHER _____

SAMPLED BY (PRINT): Betty Credit	SAMPLE DATE/TIME: 5/16/12 4:30 pm	RELINQUISHED BY: Betty Credit	DATE/TIME: 5/17/12 10:30 am
RECEIVED BY:	DATE/TIME:	RELINQUISHED BY:	DATE/TIME:
RECEIVED BY (LAB): <i>[Signature]</i>	DATE/TIME: 5/17/12 1034	PROCESSED AND VERIFIED BY: <i>[Signature]</i>	DATE/TIME: 5/17/12 1034
LOGGED IN BY: <i>[Signature]</i>	DATE/TIME: 5/17/12 1034	CARRIER: _____	COOLER TEMPERATURE: 10.6